

Certificate of Compliance

Certificate: 1088398

Master Contract: 204767

Project: 1088398

Date Issued: January 25, 2001

Issued To: Camille Bauer AG
Aargaustrasse 7
CH - 5610 Wohlen
Switzerland
Attention: Mr. Volker Pohl or Mr. Randon


*The products listed below are eligible to bear the CSA Mark shown,
with adjacent indicator "C" and "US".*



Issued by:


Eknath Desai, C.E.T.

Authorized by:


Nick Alfano
Operations Manager

PRODUCTS

CLASS 8721 04 - LABORATORY ELECTRICAL EQUIPMENT
CLASS 8721 84 - ELECTRICAL EQUIPMENT FOR LABORATORY USE

The "NRTL/C" indicator adjacent to the CSA Mark signifies that the product has been evaluated to the applicable ANSI/UL and CSA Standards, for use in the U.S. and Canada. NRTL, i.e. Nationally Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.

**

The "C" and "US" indicator adjacent to the CSA Mark signifies that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S. This includes products eligible to bear the NRTL indicator. NRTL, i.e. Nationally Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognised to perform certification to U.S. Standards.



Programmable Multi-Transducer, Model DME 4XX* Series, and DMT10XX** Series, Power Supply: AC/DC 85-230V or AC/DC 24-60V, 50/60Hz, 9W, Permanently Connected, Continuous Operation, Class II, Pollution Degree 2, Installation Category II.

* Note: The **** in the model designation number(s) may be any number 0 to 9 or letter A to Z or blank denoting SELV/ELVEL secondary circuits or minor mechanical differences.

APPLICABLE REQUIREMENTS

- | | | |
|-----------------------------|---|---|
| CAN/CSA-C22.2 No. 1010.1-92 | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |
| UL Std No. 3101.1 | - | Electrical Equipment for Laboratory Use; Part 1: General Requirements |



Supplement to Certificate of Compliance

Certificate: 1088398

Master Contract: 204767

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
1088398	2001/01/25	Original Certification